

MultiPlex Controller Guide

Prior to Procedure

Step 1: Set temperature scale

Press  for Celsius

Press  for Fahrenheit

Step 2: Adjust room temperature

Press  to increase

Press  to decrease

Step 3: Set room temperature

Press  to set

During Procedure

Start cement timer

Press  to start

Cold purge cement
(3.0 cc/min - no RF)

Hold down  to purge

After Procedure

Retract plunger

Press  to retract